

Codonics® Safe Label System® **Release Notes**

Operating Software

Version 3.1.0 SLS Release Notes

Summary

Version 3.1.0 software (SW) is a software release that is only supported on Codonics Safe Label System® (SLS) Point of Care Stations (PCS) with serial numbers that begin with 142C, 143C and 144C. Version 3.1.0 SW is NOT compatible with SLS PCS with serial numbers that begin with 140C and 141C. SLS PCS does not support software downgrades to previous software versions.

NOTE: The 3.1.0 Administration Tool (AT) software must be used with the 3.1.0 SLS PCS software.

The 3.1.0 AT works simultaneously with supported 3.1.x SLS PCSs.

This document provides information about New Features (i.e., Adaptive Updates), Product Improvements (i.e., Perfective Updates) and Product Corrections (i.e., Corrective Updates) in 3.0.0 SLS PCS software.

WARNING: The use of label colors is intended only as an aid in the identification of drug groups and does not absolve the user from the duty of reading the label and correctly identifying the drug prior to use.

New Features

- ◆ The SLS Operating System (OS) was updated. A technology refresh was completed that updated the OS to SLS OS 4. Refer to Codonics IT Resource Guide 900-653-013 for details.

Product Improvements

- ◆ The operating software includes security updates. Refer to Codonics IT Resource Guide 900-653-013 for details.
- ◆ The time zone database was updated to the latest available.

Product Corrections

- ◆ Previously there were some occurrences where a SmartDrive was not found on startup. This issue was corrected.

- ◆ Previously if logs were near a file size limit, during download of the logs, it would miss a log file. This would require re-downloading the logs to troubleshoot. This issue was corrected.

Known Common Issues

This section details common issues with SLS PCSs that are likely to be experienced by most users. These issues are specific to the current release; previous issues can be found in previous Release Notes. None of the issues included in the document are considered to render the product unsafe or ineffective for clinical use.

- ◆ **Pressing the Load button does not always cut labels on SLS PCS.** When pressing the Load button, the SLS will feed media forward but does not cut it when the door is closed. Once a subsequent label is printed, the unit will eject the printed label attached to a blank label from pressing the Load button. When the Load button is pressed on SLS PCSs with serial numbers 140C and 141C, the unit will feed media forward. When the door is closed, it will cut a label.

Known Uncommon Issues

This section details uncommon issues with all SLS PCSs that are unlikely to be experienced by most users. These issues are specific to the current release; previous issues can be found in previous Release Notes. None of the issues included in the document are considered to render the product unsafe or ineffective for clinical use.

- ◆ **The screen saver does not dismiss when touching the screen after setting date/time forward.** Plug a keyboard into the unit and press a key to dismiss the screensaver; the screen's operation will return to normal.
- ◆ **SLS PCS does not notify the user if a drug container's data matrix barcode has an invalid check digit.** All drug container's have to be verified in the AT and/or SLS PCS. Make sure you carefully review all the verification information before verifying the entry and printing a label.

Technical Support

If problems occur during software installation or operation, contact Codonics Technical Support at any time.

Phone: +1.440.243.1198

Email: support@codonics.com

Website: www.codonics.com

Get it all with just one call
1-800-444-1198

All registered and unregistered trademarks are the property of their respective owners. Specifications subject to change without notice.
Patents: www.codonics.com/ip/patents.

Copyright © 2010-2023 Codonics, Inc. Printed in the U.S.A. Part No. 901-243-041.02



17991 Englewood Drive
Cleveland, OH 44130 USA
(440) 243-1198
(440) 243-1334 Fax
Email info@codonics.com
www.codonics.com

Codonics Trading Co, Ltd.
317 Xianxia Rd. Building B
Unit 1412
Changning Dist., Shanghai
P.R. China, 200051
86-21-62787701
86-21-62787719 Fax

Codonics Limited KK
AQUACITY 9F,
4-16-23, Shibaura
Minato-ku, Tokyo, 108-0023
JAPAN
81-3-5730-2297
81-3-5730-2295 Fax