Technical Brief Codonics[®] Safe Label System[®] Operating Software

Overview

The Codonics Safe Label System (SLS) Point of Care Station (PCS) version 3.2.0 operating software is provided on a USB flash drive in the Administration Tool kit (SLS-AT-KIT-3.2.0). In the event that an individual SLS PCS requires an operating software reload, the flash drive is provided for use only with direct support by the Codonics Technical Support

WARNING: The USB flash drive should only be used for software installation. You should not use it for placing formulary or configuration packages or any other files on it since it can cause issues when executing an installation.

Operating Software

If the Codonics Technical Support team determines that an SLS PCS requires an operating software reload, it is essential that your SLS PCS v3.2.0 software flash drive (Part Number 020-120-050) be available. Codonics Technical Support will work with you to manage the process of reloading the operating software on your SLS PCS.

CAUTION: Operating software reloading must be performed only under the direction of Codonics Technical Support. Please do not attempt this process without the assistance of Codonics.

Technical Support

Please contact Codonics Technical Support at any time.

Phone: +1.440.243.1198

Email: support@codonics.com

Website: www.codonics.com

Get it all with just one call 800.444.1198

All registered and unregistered trademarks are the property of their respective owners. Specifications subject to change without notice.

Patents: www.codonics.com/ip/patents.



17991 Englewood Drive Cleveland, OH 44130 USA +1.440.243.1198 +1.440.243.1334 Fax Email info@codonics.com www.codonics.com

Codonics Trading Co, Ltd. 317 Xianxia Rd. Building B Unit 1412 Changning Dist., Shanghai P.R. China, 200051 86-21-62787701 86-21-62787719 Fax

Codonics Healthcare Japan AQUACITY 9F, 4-16-23. Shibaura Minato-ku, Tokyo, 108-0023 JAPAN 81-3-5730-2297 81-3-5730-2295 Fax